Decision Memo for Positron Emission Tomography (N-13 Ammonia) for Myocardial Perfusion (CAG-00165N)

Decision Summary

CMS determines that the evidence is adequate to conclude that the use of N-13 ammonia PET for the evaluation
of myocardial perfusion is reasonable and necessary for the diagnosis or treatment of the illness or injury or to
improve the functioning of a malformed body member in the population specified.

In addition, CMS determines that the evidence is adequate to impose the following limitations:

The PET scan, whether at rest alone, or rest with stress, is performed in place of, but not in addition to, a single photon emission computed tomography (SPECT); or

The PET scan, whether at rest alone or rest with stress, is used following a SPECT that was found to be inconclusive. In these cases, the PET scan must have been considered necessary in order to determine what medical or surgical intervention is required to treat the patient. (For purposes of this requirement, an inconclusive test is a test(s) whose results are equivocal, technically uninterpretable, or discordant with a patient's other clinical data and must be documented in the beneficiary's file.

Therefore, we intend to issue a positive national coverage determination announcing the addition of coverage for N-13 ammonia PET for the evaluation of myocardial perfusion.

Back to Top

Decision Memo

This decision memorandum does not constitute a national coverage determination (NCD). It states CMS's intent to issue an NCD. Prior to any new or modified policy taking effect, CMS must first issue a manual instruction giving specific directions to our claims-processing contractors. That manual issuance, which includes an effective date, is the NCD. If appropriate, the Agency must also change billing and claims processing systems and issue related instructions to allow for payment. The NCD will be published in the Medicare Coverage Issues Manual. Policy changes become effective as of the date listed in the transmittal that announces the Coverage Issues Manual revision.

TO: Administrative File CAG #00165N

FROM:

Steve E. Phurrough, MD, MPA Acting Director, Coverage and Analysis Group

Anthony Norris Health Insurance Specialist, Coverage and Analysis Group


SUBJECT: N-13 Ammonia and Positron Emission Tomography for Myocardial Perfusion


DATE: April 16, 2003

This memo serves four purposes: (1) briefly outlines the assessment of coronary artery disease and congestive heart failure using N-13 ammonia and positron emission tomography; (2) reviews the history of Medicare's coverage process on the use of PET for myocardial perfusion; (3) presents and analyzes the relevant scientific literature assessing the clinical effectiveness of N-13 ammonia PET and comparing it to Rubidium-82 (Rb-82) for the detection of myocardial perfusion; (4) announces our intention to issue an NCD revising CIM 50-36 and delineates the reasons for making a positive coverage determination for the use of N-13 ammonia PET for assessing myocardial perfusion.

Clinical Background

Cardiovascular disease (CVD) is a broad term encompassing several conditions, such as, coronary artery disease (CAD), congestive heart failure (CHF), and stroke. CVD is the number one cause of death in the United States, claiming 949,619 lives in 1998. Of that total, 459,841 deaths were attributable to CAD. Eighty-five percent of the people who die of CAD are age 65 or older and, thus, may be Medicare beneficiaries.





Access to nuclear imaging is a crucial issue for Medicare beneficiaries. SPECT is widely available throughout the US. PET imaging is still in its early diffusion phase and many communities lack this technology. In addition, Rb-82 has a different manufacturing process from FDG and N-13 ammonia. Rb-82 is generator-produced while the other two are produced in a cyclotron. All three isotopes have a relatively short half-life, necessitating local production. Therefore, using Rb-82 and FDG requires a community to support both technologies.

History of Medicare's Coverage on PET for Myocardial Perfusion


On March 14, 1995, CMS began covering PET for myocardial perfusion using the radiopharmaceutical Rb-82. In the evaluation of myocardial perfusion, FDG PET is covered as a primary or initial diagnostic study, or following an inconclusive SPECT, prior to revascularization. SPECT may not be used following an inconclusive PET.

In addition, in 1998, CMS began covering the use of FDG PET or SPECT for the evaluation of myocardial viability⁵. It also can be used as an initial study or following a SPECT that was inconclusive.

Timeline of Recent Activity

September CMS elected to begin a national coverage determination process following the receipt of a formal request from R. Edward Coleman, MD, Director, Division of Nuclear Medicine, Duke University Medical Center, "for a national coverage decision for N-13 ammonia for PET imaging of the myocardium under rest and/or pharmacologic stress conditions to evaluate myocardial perfusion in patients with suspected or existing coronary disease."

Food and Drug Administration (FDA) Status



- 1. A large, systematic review of the literature performed by the FDA, dated February 18, 1999, used as the basis for FDA approval of N-13 ammonia.
- 2. A literature review with evidence tables submitted as part of the formal request package by R. Edward Coleman, $MD._{0}^{10}$
- An OVID MEDLINE search.

FDA Systematic Review

The FDA systematic review is the most comprehensive review identified. Search criteria included: studies published from January 1990 to July 1, 1998 identified as human clinical trials with N-13 ammonia PET, written in English, found by searching on-line databases of Medline, Cancerlit, Derwent Drug File, Biosis Preview, International Pharmacology Abstracts and Embase. Review articles on N-13 ammonia PET imaging were identified using the same criteria in the Cochrane Database for Systemic Reviews and Cochrane Controlled Trials Register. Finally, FDA solicited references from the PET community on N-13 ammonia PET from any time period published in peer-reviewed journals.

In all, 17 articles were included in the review of effectiveness. Two articles were reviewed as being adequate and well-controlled published studies, 3 were reviewed as controlled studies that were supportive, and 9 were reviewed as other published studies that had supporting information.

Gould (1986) and Demer (1989) were the two adequate, well-controlled studies that allowed assessment of the effectiveness by meeting the following criteria:

- There was comparison between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion perfusion perfusion perfusion between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion perfusion between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion perfusion between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion between N-13 ammonia PET and an accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 ammonia PET and accepted "truth" standard of coronary perfusion between N-13 am
- The study population was prospectively enrolled;
- The entry criteria defined the target clinical population in which N-13 ammonia PET was intended to be used:
- There were clearly defined endpoints;
- Detailed data on findings were presented; and
- There were procedures to minimize interpretation bias, such as masking (also referred to as blinding) and randomization, when flow measurement was not performed quantitatively.

The first article, Gould (1986), is the initial report of findings and the second article, Demer (1989), is the comprehensive report that includes data from the first study.

The strengths of these studies include:

- The proper population for PET imaging was studied; i.e., those who had indications of possible CAD and were undergoing cardiac catheterization.
- It was a prospective collection of patients.
- There were subjects with normal coronary anatomy included in the selection.
- Although small in sample size, the study had the appropriate enrollment criteria to encompass a board spectrum of the disease, patients with possible CAD.
- PET imaging was compared to an accepted standard, cardiac angiography.
- The interpreting physicians were masked and read images of the same patient three times.


Weakness in these studies included:

- The first study lacked information on inter-reader variability, whether images were read independently, the frequency of disagreement of PET interpretations, and dispute resolution. This was corrected in the second study.
- Results were not stratified by radiopharmaceutical (Rb-82 versus N-13 ammonia). However, the study did mention, "There was no difference in sensitivity and specificity between the two tracers in this small number of patients." This implies that data was analyzed to support this statement, but was not published.
- There were a small number of subjects.

In Demer (1989), 111 patients received PET perfusion imaging with N-13 ammonia and the remaining 82 patients with Rb-82. Thus, the two radiopharmaceuticals were not compared to each other in the same patients. Therefore, one cannot compare the two radiopharmaceuticals directly but must draw conclusions from the performance characteristics determined separately. Two independent readers, who were blinded to clinical data, visually interpreted rest and stress images. Arteriographic stenosis flow reserve (SFR) was compared to PET defect severity, with 105/111 (5 false negative and 1 false positive) N-13 ammonia images and 79/82 (2 false negative and 1 false positive) Rb-82 images providing findings consistent with arteriographic results. These ratios were not significantly different (p=0.73), and qualitative similarities were also noted between both tracer images. Such data suggest equivalent performance between the tracers.

The FDA report concluded that, overall, these studies were well-designed and provided results supporting similar performance of N-13 ammonia PET to Rb-82 PET for assessing coronary perfusion.

Three other articles were reviewed by the FDA that they felt provided supporting information: Schelbert (1982), DeCarli & Sherman (1994), and Gerwitz (1994).



Requestor's Review

The requestor submitted a large volume of literature with evidence tables and a summary. This was not a systematic review and the strengths and weaknesses of the studies were not assessed. Only one article met the criteria of testing against a gold standard, Schelbert (1982). The remainder evaluated differing hypotheses for the use of N-13 ammonia PET. The summary specified an overall approximate sensitivity of 90% and a specificity of 90% though the method of calculating them was not given.

CMS Evidentiary Review

CMS performed a search for English language and human subjects only, using the keywords "N-13 Ammonia" and "Rb-82" and "SPECT" and "Myocardial Perfusion" for publication dates after 1990. Review of reference lists in these articles yielded additional articles, including those predating 1990. Review articles, articles without a "gold standard" for comparison, and articles with a hypothesis not assessing perfusion in CAD were excluded.

Only five articles met our criteria. All predated 1990. The acceptance of N-13 ammonia as a nuclear imaging tracer in clinical practice before 1990 has resulted in subsequent studies addressing other issues to include:


- Ability of N-13 ammonia PET to predict outcomes in high risk patients;
- Predicting recovery from acute myocardial injury;
- Use in diseases other than CAD; e.g. hypertrophic cardiomyopathy;
- Differentiating ischemia and infarction acutely; and
- Quantifying myocardial flow by region.

Studies that directly addressed the sensitivity and specificity of N-13 ammonia PET as a diagnostic test for myocardial perfusion are summarized in the evidence table below. Angiographic findings were used as the gold standard in each of these studies. Sensitivities ranged from 88-97% and specificities from 90-100%.

Author/Year	Subjects	Gold Standard	Sensitivity	Specificity	Stress Source
Schelbert (1982) <u>11</u>	32 with documented coronary artery disease (CAD) and 13 normal	Angiography	31/32 (97%) with abnormal PET:	13/13 (100%)	Dipyridamole-induced coronary hyperemia

Author/Year	Subjects	Gold Standard	Sensitivity	Specificity	Stress Source
	volunteers		One false negative		
Yonekura (1987) <u>12</u>	40 patients with CAD by angiography; 26 with prior MI	Angiography	After exercise, 37/38 (97%) of patients with CAD.	12/12 (100%)	Exercise
	All 40 had rest PET studies; only 38 had exercise studies (2 became unstable).		67/75 (89%) of all stenosed vessels		
	20 normal volunteers; 9 with atypical chest pain and normal angiography and 11 w/o clinical, ECG or physical evidence of CAD. 8 were used to establish minimum limits for normal values. Only 12 were studied.				
Tamaki (1988) <u>13</u>	51 patients referred for coronary angiography without further elaboration of clinical presentation, other than 38 had history of myocardial infarction	Angiography	47/48 (97%) for patients with abnormal angiograms.	3/3 (100%) for angiographically normal patients	Exercise
			80/91 (88%) of all stenosed vessels.	56/62 arteries (90%) for vessels without disease	

Schelbert (1982) has been discussed above.
Yonekura (1987) enrolled 60 patients in a study to assess the accuracy of N-13 ammonia PET in diagnosing CAD as compared to an angiographic gold standard. Forty patients (36 men, 4 women; average 52, range 34 – 72) with angiographically demonstrated disease were selected and underwent rest and exercise PET studies (2 patients with 3 vessel disease became unstable and did not progress to the exercise study). The 20 normal volunteers (18 men, 2 women; average age 45, range 25 – 65) included 9 who had angiography for atypical chest pain. The remaining 11 had no evidence of CAD by clinical and ECG exam. They did not undergo angiography. Eight of the 20 normal subjects studies were limited to determination of minimum limits of the normal values.
The strength of this study is the comparison of diseased and normal subjects using an objective gold standard—angiography. Weaknesses include a lack of discussion of verification of the PET results and a large portion of the "normal" population did not have their lack of disease verified by the gold standard.
The results demonstrated a sensitivity of 97% in identifying patients with CAD and 89% in identifying diseased vessels. If one assumes that all the normal patients were disease free, then the sensitivity for identifying patients without CAD was 100%.
Tamaki (1988) studied 51 patients who were referred for coronary angiography and performed both stress thallium SPECT and stress N-13 ammonia PET. Gender was not specified. The age ranged from 34 to 71 with a mean of 56.1. Thirty-eight patients had a history of myocardial infarction. The angiograms and nuclear images were all independently interpreted by three experienced observers without knowledge of patient conditions or results of other tests. Though this was a comparison of N-13 ammonia PET to thallium SPECT, the results of the N -13 ammonia PET data are illustrative.
This is a well-designed study. Appropriate patients are selected, a gold standard is used, and masking appears appropriate. A weakness is the lack of discussion of patient selection.
The results of the N-13 ammonia PET data demonstrates a sensitivity of 98% in identifying patients with CAD and a sensitivity of 88% in identifying diseased vessels. Three patients had normal angiograms and all three had a negative N-13 ammonia PET for a specificity of 100% for patients with disease. Specificity was 90% for diseased vessels.



National coverage determinations (NCDs) are determinations by the Secretary with respect to whether or not a particular item or service is covered nationally under title XVIII of the Social Security Act \S 1869(f)(1)(B). In order to be covered by Medicare, an item or service must fall within one or more benefit categories contained within Part A or Part B, and must not be otherwise excluded from coverage. Moreover, with limited exceptions, the expenses incurred for items or services must be "reasonable and necessary for the diagnosis or treatment of illness or injury or to improve the functioning of a malformed body member." \S 1862(a)(1)(A).


CMS has issued regulations pertaining to the coverage of diagnostic tests under the Part B program. Those rules provide that, except for a few exceptions, diagnostic tests must be ordered by the physician who treats the beneficiary for a specific medical problem and the physician must use the results in the management of the beneficiary's specific medical problem (42 C.F.R. § 410.32). In general, tests not ordered by the physician who is treating the beneficiary are not reasonable and necessary. See also 42 C.F.R. § 411.15(k)(1).


CMS believes that the evidence discussed above is sufficient to determine if N-13 ammonia PET is reasonable and necessary. The studies discussed above demonstrate relatively high values for sensitivity (88-97%) and specificity (90-100%). These studies are well designed and use the appropriate gold standard of angiographic findings. Also, support from the two major specialty societies, as well as testimony from key experts who attest to the fact that N-13 ammonia PET has achieved mainstream use in the practice of nuclear cardiology, provide additional evidentiary support.


Accuracy in assessing myocardial perfusion and matching it to myocardial viability studies provides crucial information to physicians attempting to select appropriate patients for revascularization. Revascularizing stenotic vessels that perfuse nonviable myocardium is not felt to be beneficial. Thus, this information allows physicians to alter patient management decisions. CMS believes that the evidence is adequate to conclude that the use of N-13 ammonia PET in clinical practice has the potential to improve net health outcomes through changes in patient management..


Although the Demer et al. study, which compares N-13 ammonia and Rb-82, fails to test both tracers head-to-head in the same population, it was sufficiently well crafted to demonstrate that both tracers have high reliability and validity in assessing myocardial perfusion. CMS believes the evidence is adequate to conclude that both N-13 ammonia PET and Rb-82 PET are of equal accuracy and reliability when used in the evaluation of myocardial perfusion.













Houn Florence, Castillo Florence. Medical and statistical review of N-13 ammonia positron emission tomography. February 18, 1999.
(http://www.fda.gov/cder/regulatory/pet/ammonfinal.htm).
Hutchins G, Schwaiger M, Rosenspire K, Krivokapich J, Schelbert H, Kuhl D. Noninvasive quantification of regiona blood flow in the human heart using N-13 ammonia and dynamic positron emission tomographic imaging. <i>Journa of the American College of Cardiology 1990;</i> 15(5):1032-1042.
Kambara H, Nohara R, Tamaki N, Ono S, Konishi J, Kawai C. Positron emission tomography –usefulness in assessing myocardial viability. <i>Japanese Circulation Journal</i> 1992;56:608-613.
Konishi Y, Toshihiko B, Yoshifumi O, et al. Myocardial positron tomography with N-13 ammonia in assessment of aortocoronary bypass surgery. <i>Japanese Circulation Journal</i> 1988; 52:411-416.
Kosa I, Blasini R, Schneider-Eicke J, et al. Early recovery of coronary flow reserve after stent implantation as assessed by positron emission tomography. <i>Journal of the American College of Cardiology</i> 1999;34(4):1036-41.
Krivokapich, J, Huang, S, Ratib, O, Schelbert, H. Noninvasive detection of functionally significant coronary artery stenoses with exercise and positron emission tomography. <i>American Heart Journal</i> 1991;122:202-211.
Krivokapich J, Stevenson L, Kobashigawa J, Huang SC, Schelbert H. Quantification of absolute myocardial perfusion at rest and during exercise with positron emission tomography after human cardiac transplantation. Journal of the American College of Cardiology1991;18(2):512-517.
Lancellotti P, Melon PG, de Landsheere CM, Degueldre C, Kulbertus HE, Pierard LA. The role of early measurement of nitrogen-13 ammonia uptake for predicting contractile recovery after acute myocardial infarction. <i>International Journal of Cardiac Imaging</i> 1998;14(4):261-267; discussion 269-70.





Vanoverschelde JL, Depre C, Gerber BL, et al. Time course of functional recovery after coronary artery bypass graft surgery in patients with chronic left ventricular ischemic dysfunction. *American Journal of Cardiology* 2000;85(12):1432-1439.

Yates RW, Marsden PK, Badawi RD, et al. Evaluation of myocardial perfusion using positron emission tomography in infants following a neonatal arterial switch operation. *Pediatric Cardiology* 2000;21(2):111-118.

Yonekura Y, Tamaki N, Senda M, et al. Detection of coronary artery disease with 13N-ammonia and high resolution postiron-emission computed tomography. *American Heart Journal* 1987;113(3):645-654.

Back to Top

Related Material

Other Material1 Back to Top